UNITED STATES DEPARTMENT OF COMMERCE United States Patent and Trademark Office Address: COMMISSIONER FOR PATENTS P.O. Box 1450 Alexandria, Virginia 22313-1450 www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/781,426	02/18/2004	David J. Jorgenson	P0008050.05	3045
27581 MEDTRONIC,	7590 10/29/200 INC.	8	EXAMINER	
710 MEDTRON	NIC PARKWAY NE		HOLMES, REX R	
MINNEAPOLIS, MN 55432-9924			ART UNIT	PAPER NUMBER
			3762	
			MAIL DATE	DELIVERY MODE
			10/29/2008	PAPER

Please find below and/or attached an Office communication concerning this application or proceeding.

The time period for reply, if any, is set in the attached communication.

	Application No.	Applicant(s)				
	10/781,426	JORGENSON ET AL.				
Office Action Summary	Examiner	Art Unit				
	REX HOLMES	3762				
The MAILING DATE of this communication appears on the cover sheet with the correspondence address Period for Reply						
A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION. - Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication. - If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication. - Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).						
Status						
1) Responsive to communication(s) filed on 27 Ma	arch 2008					
	action is non-final.					
<i>,</i> —	Since this application is in condition for allowance except for formal matters, prosecution as to the merits is					
	closed in accordance with the practice under <i>Ex parte Quayle</i> , 1935 C.D. 11, 453 O.G. 213.					
Disposition of Claims						
4)⊠ Claim(s) <u>17-24</u> is/are pending in the application.						
	4a) Of the above claim(s) is/are withdrawn from consideration.					
5) Claim(s) is/are allowed.						
6)⊠ Claim(s) <u>17-24</u> is/are rejected.						
7) Claim(s) is/are objected to.						
8) Claim(s) are subject to restriction and/or	election requirement.					
Application Papers						
9) The specification is objected to by the Examiner.						
10) The drawing(s) filed on is/are: a) accepted or b) objected to by the Examiner.						
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).						
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).						
11) The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.						
Priority under 35 U.S.C. § 119						
 12) Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f). a) All b) Some * c) None of: 1. Certified copies of the priority documents have been received. 2. Certified copies of the priority documents have been received in Application No. 3. Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)). * See the attached detailed Office action for a list of the certified copies not received. 						
Attachment(s) 1) X Notice of References Cited (PTO-892)	4) 🔲 Interview Summery	(PTO-413)				
2) Notice of Draftsperson's Patent Drawing Review (PTO-948) Paper No(s)/Mail Date						
3) Information Disclosure Statement(s) (PTO/SB/08) 5) Notice of Informal Patent Application						
Paper No(s)/Mail Date 6) Other:						

Art Unit: 3762

DETAILED ACTION

Claim Rejections - 35 USC § 102

1. The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless -

- (b) the invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.
- 2. Claims 17-19 and 22-23 are rejected under 35 U.S.C. 102(b) as being anticipated by Schuelke et al. (U.S. Pat. 5,755,742 hereinafter "Schuelke").
- 3. Regarding claims 17 and 22-23, Schuelke discloses a system that includes a implantable device that allows for sensing in multiple pathways (e.g. Col. 15, II. 30-40), that also senses a p-wave amplitude (e.g. Col. 10, II. 15-28), and further processes the data for each pathway to determine if there is a lead related event (e.g. Col. 16, II. 43-58; see also Abstract).
- 4. Regarding claim 18, Schuelke discloses that the system further alerts a user if a lead condition occurs (e.g. Col. 15, II. 9-25).
- 5. Regarding claim 19, Schuelke discloses that the system detects if it is a lead conductor or insulation issue (e.g. Col. 15, II. 30-40).

Claim Rejections - 35 USC § 103

- 6. The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:
 - (a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negatived by the manner in which the invention was made.

Art Unit: 3762

7. This application currently names joint inventors. In considering patentability of the claims under 35 U.S.C. 103(a), the examiner presumes that the subject matter of the various claims was commonly owned at the time any inventions covered therein were made absent any evidence to the contrary. Applicant is advised of the obligation under 37 CFR 1.56 to point out the inventor and invention dates of each claim that was not commonly owned at the time a later invention was made in order for the examiner to consider the applicability of 35 U.S.C. 103(c) and potential 35 U.S.C. 102(e), (f) or (g) prior art under 35 U.S.C. 103(a).

- 8. Claim 20 is rejected under 35 U.S.C. 103(a) as being unpatentable over Schuelke as applied to claims 17-19 above, and further in view of Paul et al. (U.S. pat. 5,814,088 hereinafter "Paul").
- 9. Schuelke discloses the claimed invention except for the lead status event being a biological interface issue. However, Paul teaches that it is known to use a biological interface issue as a lead status event as set forth in Col. 11, II. 20-23 to provide a further reason why a lead failed without. It would have been obvious to one having ordinary skill in the art at the time the invention was made to modify the system as taught by Schuelke, with the biological interface issues as taught by Paul, since such a modification would provide the predictable results of a lead failure monitoring system for providing a more informative system to show loss of lead function due to both lead failure and lead dislocation.

Art Unit: 3762

Allowable Subject Matter

10. Claims 24 objected to as being dependent upon a rejected base claim, but would be allowable if rewritten in independent form including all of the limitations of the base claim and any intervening claims.

Response to Arguments

11. Applicant's arguments with respect to claims 17-24 have been considered but are moot in view of the new ground(s) of rejection.

Conclusion

Any inquiry concerning this communication or earlier communications from the examiner should be directed to REX HOLMES whose telephone number is (571)272-8827. The examiner can normally be reached on M-F 8:00 - 5:00.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Angela Sykes can be reached on 571-272-4955. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Art Unit: 3762

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see http://pair-direct.uspto.gov. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

/R. H./ Examiner, Art Unit 3762 /George R Evanisko/ Primary Examiner, Art Unit 3762